

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050719\
Data File : VL033678.D
Acq On : 7 May 2019 12:22
Operator : SY/AP
Sample : VL0507ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0507ABL01

Quant Time: May 08 03:20:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050619AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon May 06 15:19:01 2019
QLast Update : Mon May 06 15:19:01 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	911638	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	2112168	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2073123	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1801375	10.60	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.00%

Target Compounds					Ovalue
15) Acetone	3.92	43	21846	0.161 ppbv	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050719\
Data File : VL033678.D
Acq On : 7 May 2019 12:22
Operator : SY/AP
Sample : VL0507ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0507ABL01

Quant Time: May 08 03:20:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050619AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon May 06 15:19:01 2019
QLast Update : Mon May 06 15:19:01 2019
Response via : Initial Calibration

