

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050720\
 Data File : VL034974.D
 Acq On : 7 May 2020 19:11
 Operator : SY/AP
 Sample : L2524-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IIA2DL

Manual Integrations
 APPROVED

MMDadoda
 5/8/2020 5:22:43 PM

Quant Time: May 08 05:37:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 0
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1026900	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2357576	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2203229	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1823642	10.23	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	136564	8.135	ppbv	95
15) Acetone	3.84	43	194540	1.433	ppbv	98
19) Isopropyl Alcohol	3.95	45	1191009	13.155	ppbv	99
20) Methylene Chloride	4.33	84	17490m	0.389	ppbv	
26) Hexane	5.68	57	14006	0.142	ppbv #	44
34) 2-Butanone	5.25	43	35496	0.242	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050720\
Data File : VL034974.D
Acq On : 7 May 2020 19:11
Operator : SY/AP
Sample : L2524-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IIA2DL

Manual Integrations
APPROVED

MMDadoda
5/8/2020 5:22:43 PM

Quant Time: May 08 05:37:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

