

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL050721\
 Data File : VL036914.D
 Acq On : 7 May 2021 11:48
 Operator : SY/AP
 Sample : VL0507ABL01
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0507ABL01

Manual Integrations
 APPROVED

MMDadoda
 5/10/2021 3:34:25 PM

Quant Time: May 08 18:53:23 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LTue Apr 27 16:56:35 2021
 QLast Update : Tue Apr 27 16:56:35 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1306057	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	7.18	114	3078497	10.00	ppbv	0.01
55) Chlorobenzene-d5	12.09	117	2608595	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2046056	9.77	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.89	43	36348m	0.295	ppbv	
20) Methylene Chloride	4.35	84	10437m	0.177	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL050721\
Data File : VL036914.D
Acq On : 7 May 2021 11:48
Operator : SY/AP
Sample : VL0507ABL01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0507ABL01

Manual Integrations
APPROVED

MMDadoda
5/10/2021 3:34:25 PM

Quant Time: May 08 18:53:23 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL042721AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Apr
QLast Update : Tue Apr 27 16:56:35 2021
Response via : Initial Calibration

