

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050725\
 Data File : VL042516.D
 Acq On : 07 May 2025 13:50
 Operator : SY/MD
 Sample : MDL2 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL2 0.03

Quant Time: May 08 01:35:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 05/08/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	74074	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	183542	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	155750	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	138229	9.573	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.700%
Target Compounds						
						Qvalue
32) 1,1,1-Trichloroethane	3.373	97	646m	0.034	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.976	83	524m	0.035	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050725\
Data File : VL042516.D
Acq On : 07 May 2025 13:50
Operator : SY/MD
Sample : MDL2 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL2 0.03

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/08/2025
Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 08 01:35:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration

