

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050725\
 Data File : VL04252EX3.D
 Acq On : 07 May 2025 18:32
 Operator : SY/MD
 Sample : CAN 10198 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10198 40X

Quant Time: May 08 08:14:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 02 03:16:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	76470	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	179584	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.891	117	153592	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	136877	9.612	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.100%
Target Compounds						
						Qvalue
15) Acetone	1.839	43	65037	4.659	ppbv	97
20) Methylene Chloride	2.065	84	505193	120.377	ppbv	99
26) Hexane	2.829	57	384900	39.528	ppbv #	38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050725\
Data File : VL04252EX3.D
Acq On : 07 May 2025 18:32
Operator : SY/MD
Sample : CAN 10198 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10198 40X

Quant Time: May 08 08:14:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL050125AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 02 03:16:31 2025
Response via : Initial Calibration

