

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050924\
 Data File : VL041249.D
 Acq On : 9 May 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 10 01:11:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 10 01:10:17 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	1.353	1.506	-11.3	87	0.00
3	Chlorodifluoromethane	1.519	1.638	-7.8	86	0.00
4	Chloromethane	0.596	0.620	-4.0	82	0.00
5 T	Vinyl Chloride	0.599	0.635	-6.0	82	0.00
6 T	Bromomethane	0.218	0.234	-7.3	83	0.00
7	Chloroethane	0.211	0.216	-2.4	80	0.00
8 T	Dichlorotetrafluoroethane	1.400	1.540	-10.0	86	0.00
9 T	Propene	0.594	0.559	5.9	75	0.00
10 T	Heptane	1.440	1.438	0.1	79	0.00
11 T	Trichlorofluoromethane	1.375	1.578	-14.8	89	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.017	1.106	-8.8	84	0.00
13	Ethanol	0.133	0.146	-9.8	80	0.00
14 T	Bromoethene	0.420	0.456	-8.6	84	0.00
15 T	Acetone	1.180	1.180	0.0	86	0.00
16 T	1,3-Butadiene	0.596	0.645	-8.2	82	0.00
17	tert-Butyl alcohol	1.273	1.146	10.0	64	0.00
18 T	1,1-Dichloroethene	0.453	0.492	-8.6	83	0.00
19 T	Isopropyl Alcohol	0.774	0.830	-7.2	76	0.00
20 T	Methylene Chloride	0.410	0.410	0.0	81	0.00
21 T	Allyl Chloride	0.770	0.780	-1.3	79	0.00
22 T	trans-1,2-Dichloroethene	0.471	0.532	-13.0	86	0.00
23 T	Vinyl Acetate	1.300	1.362	-4.8	79	0.00
24 T	1,1-Dichloroethane	1.011	1.095	-8.3	86	0.00
25 T	Ethyl Acetate	2.356	2.407	-2.2	80	0.00
26 T	Hexane	1.051	1.050	0.1	80	0.00
27 T	Carbon Disulfide	1.244	1.338	-7.6	81	0.00
28 T	Methyl tert-Butyl Ether	0.837	0.806	3.7	74	0.00
29 T	Chloroform	1.450	1.640	-13.1	88	0.00
30 T	Cyclohexane	0.832	0.876	-5.3	80	0.00
31 T	cis-1,2-Dichloroethene	0.984	1.050	-6.7	82	0.00
32 T	1,1,1-Trichloroethane	1.447	1.631	-12.7	87	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
34 T	2-Butanone	0.509	0.490	3.7	86	0.00
35 T	Carbon Tetrachloride	0.561	0.631	-12.5	89	0.00
36 T	Benzene	0.839	0.839	0.0	82	0.00
37 T	1,2-Dichloroethane	0.438	0.483	-10.3	89	0.00
38 T	Trichloroethene	0.313	0.326	-4.2	90	0.00
39 T	1,2-Dichloropropane	0.323	0.320	0.9	80	0.00
40 T	1,4-Dioxane	0.116	0.124	-6.9	84	0.00
41 T	Tetrahydrofuran	0.368	0.344	6.5	76	0.00
42 T	Bromodichloromethane	0.628	0.694	-10.5	87	0.00
43	Methyl Methacrylate	0.339	0.334	1.5	78	0.00
44 T	2,2,4-Trimethylpentane	1.447	1.410	2.6	81	0.00
45 T	t-1,3-Dichloropropene	0.246	0.259	-5.3	76	0.00
46 T	cis-1,3-Dichloropropene	0.301	0.309	-2.7	80	0.00
47 T	1,1,2-Trichloroethane	0.324	0.344	-6.2	87	0.00
48 T	Dibromochloromethane	0.516	0.584	-13.2	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL050924\
 Data File : VL041249.D
 Acq On : 9 May 2024 8:46
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 10 01:11:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 10 01:10:17 2024
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.439	0.486	-10.7	84	0.00
50 T	4-Methyl-2-Pentanone	0.913	0.885	3.1	81	0.00
51 T	2-Hexanone	0.678	0.651	4.0	81	0.00
52 T	Tetrachloroethene	0.295	0.296	-0.3	83	0.00
53 T	Toluene	0.973	0.982	-0.9	83	0.00
54 T	1,2-Dibromoethane	0.438	0.483	-10.3	88	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
56	1,1,1,2-Tetrachloroethane	0.405	0.425	-4.9	88	0.00
57 T	Chlorobenzene	0.732	0.728	0.5	87	0.00
58 T	Ethyl Benzene	1.371	1.310	4.4	85	0.00
59 T	m/p-Xylene	1.152	1.114	3.3	87	0.00
60 T	o-Xylene	1.080	1.050	2.8	87	0.00
61 T	Styrene	0.571	0.576	-0.9	86	0.00
62	Isopropylbenzene	1.616	1.567	3.0	88	0.00
63 T	1,1,2,2-Tetrachloroethane	0.736	0.731	0.7	87	0.00
64	n-propylbenzene	0.403	0.410	-1.7	88	0.00
65	tert-Butylbenzene	1.447	1.381	4.6	89	0.00
66 T	Benzyl Chloride	0.164	0.119	27.4	55	0.00
67	sec-Butylbenzene	2.037	1.960	3.8	89	0.00
68 S	1-Bromo-4-Fluorobenzene	0.755	0.729	3.4	79	0.00
69	p-Isopropyltoluene	1.679	1.631	2.9	89	0.00
70	n-Butylbenzene	1.753	1.705	2.7	89	0.00
71	2-Chlorotoluene	1.230	1.203	2.2	87	0.00
72 T	4-Ethyltoluene	1.224	1.248	-2.0	88	0.00
73 T	1,3,5-Trimethylbenzene	1.107	1.106	0.1	88	0.00
74 T	1,2,4-Trimethylbenzene	1.209	1.193	1.3	89	0.00
75 T	1,3-Dichlorobenzene	0.693	0.714	-3.0	90	0.00
76 T	1,4-Dichlorobenzene	0.667	0.700	-4.9	89	0.00
77 T	1,2-Dichlorobenzene	0.693	0.685	1.2	90	0.00
78 T	Hexachloro-1,3-Butadiene	0.616	0.489	20.6	80	0.00
79 T	Naphthalene	1.151	1.194	-3.7	89	0.00
80 T	Naphthalene,2-methyl-	0.463	0.356	23.1	77	0.00
81 T	1,2,4-Trichlorobenzene	0.608	0.575	5.4	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0