

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051122\
 Data File : VL039031.D
 Acq On : 11 May 2022 22:24
 Operator : SY/AP
 Sample : N2753-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/15/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 12 02:01:42 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LThu May 12 02:01:42 2022
 QLast Update : Thu May 05 10:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.645	49	985245	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	7.150	114	2718249	10.000	ppbv	-0.02
55) Chlorobenzene-d5	12.056	117	2322510	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.381	95	1681051	9.140	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	91.400%
Target Compounds						
9) Propene	3.002	41	24465	0.458	ppbv	96
10) Heptane	8.095	43	53492	0.408	ppbv	97
13) Ethanol	3.594	45	61978m	16.404	ppbv	
15) Acetone	3.838	43	227934	2.453	ppbv #	83
20) Methylene Chloride	4.327	84	25290	0.722	ppbv	90
26) Hexane	5.668	57	335715	3.397	ppbv #	18
34) 2-Butanone	5.227	43	306435	2.210	ppbv	95
52) Tetrachloroethene	11.192	164	29721m	0.358	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051122\
Data File : VL039031.D
Acq On : 11 May 2022 22:24
Operator : SY/AP
Sample : N2753-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/15/2022
Supervised By : Mahesh Dadoda 05/16/2022

Quant Time: May 12 02:01:42 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050522AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May

QLast Update : Thu May 05 10:12:15 2022

Response via : Initial Calibration

