

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051123\
 Data File : VL040414.D
 Acq On : 12 May 2023 9:01
 Operator : SY/MD
 Sample : 02761-01DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-03-20230511DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/14/2023
 Supervised By :Mahesh Dadoda 05/15/2023

Quant Time: May 12 15:57:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Thu Apr 13 14:37:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1148766	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	6.639	114	3008386	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.458	117	2650417	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.773	95	2065215	9.423	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.200%
Target Compounds						
9) Propene	2.674	41	68068	1.148	ppbv	97
11) Trichlorofluoromethane	3.542	101	12845m	0.061	ppbv	
13) Ethanol	3.234	45	46977	9.489	ppbv	91
15) Acetone	3.449	43	810936	4.984	ppbv #	68
34) 2-Butanone	4.774	43	588494	3.628	ppbv	98
53) Toluene	9.214	91	22104m	0.085	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051123\
Data File : VL040414.D
Acq On : 12 May 2023 9:01
Operator : SY/MD
Sample : 02761-01DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-03-20230511DL

Quant Time: May 12 15:57:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL041323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Apr 13 14:37:02 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/14/2023
Supervised By : Mahesh Dadoda 05/15/2023

