

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051225\
 Data File : VL042530.D
 Acq On : 12 May 2025 11:49
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/14/2025
 Supervised By :Mahesh Dadoda 05/14/2025

Quant Time: May 12 23:49:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon May 12 23:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.780	49	96870	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.949	114	261294	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.872	117	223954	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.370	95	190537	9.778	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.576	62	229	0.035	ppbv #	4
32) 1,1,1-Trichloroethane	3.360	97	842m	0.036	ppbv	
35) Carbon Tetrachloride	3.761	117	785m	0.033	ppbv	
38) Trichloroethene	4.531	130	361m	0.036	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.959	83	786m	0.037	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051225\
Data File : VL042530.D
Acq On : 12 May 2025 11:49
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/14/2025
Supervised By :Mahesh Dadoda 05/14/2025

Quant Time: May 12 23:49:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon May 12 23:42:45 2025
Response via : Initial Calibration

