


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL051419\
 Data File : VL033684.D
 Acq On : 14 May 2019 9:25
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050619AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon May 06 15:19:01 2019
 Last Update : Mon May 06 15:19:01 2019



AutoFind: Scans 3616, 3617, 3618; Background Corrected with Scan 3591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.8	194389	PASS
75	95	30	66	57.5	469696	PASS
95	95	100	100	100.0	816448	PASS
96	95	5	9	6.6	54018	PASS
173	174	0.00	2	1.0	5301	PASS
174	95	50	120	67.0	546965	PASS
175	174	4	9	7.1	38709	PASS
176	174	93	101	97.2	531520	PASS
177	176	5	9	6.3	33266	PASS