

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051420\
 Data File : VL035019.D
 Acq On : 14 May 2020 13:54
 Operator : SY/AP
 Sample : VL0514ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0514ABL01

Manual Integrations
 APPROVED

MMDadoda
 5/18/2020 1:34:38 PM

Quant Time: May 15 01:49:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 0
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1011446	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2325398	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.10	117	2213007	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	1882488	10.51	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.83	43	35110	0.263	ppbv	# 57
20) Methylene Chloride	4.31	84	8256m	0.187	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051420\
Data File : VL035019.D
Acq On : 14 May 2020 13:54
Operator : SY/AP
Sample : VL0514ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0514ABL01

Manual Integrations
APPROVED

MMDadoda
5/18/2020 1:34:38 PM

Quant Time: May 15 01:49:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

