

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051420\
 Data File : VL035022.D
 Acq On : 14 May 2020 16:26
 Operator : SY/AP
 Sample : L2567-05DL DUP 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-3DL DUP

Manual Integrations
 APPROVED

MMDadoda
 5/18/2020 6:47:51 PM

Quant Time: May 18 09:02:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May 0
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1031788	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.17	114	2326564	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.10	117	2272903	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1986825	10.80	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.57	45	19036m	1.129	ppbv	
15) Acetone	3.82	43	114516m	0.840	ppbv	
19) Isopropyl Alcohol	3.94	45	132707	1.459	ppbv	# 94
53) Toluene	9.81	91	1786967	8.788	ppbv	99
59) m/p-Xylene	12.99	91	132327m	0.535	ppbv	
61) Styrene	13.55	104	91407m	0.592	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051420\
Data File : VL035022.D
Acq On : 14 May 2020 16:26
Operator : SY/AP
Sample : L2567-05DL DUP 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-3DL DUP

Manual Integrations
APPROVED

MMDadoda
5/18/2020 6:47:51 PM

Quant Time: May 18 09:02:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

