

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051420\
 Data File : VL035032.D
 Acq On : 14 May 2020 22:51
 Operator : SY/AP
 Sample : L2566-04DL2 1200X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-2DL2

Quant Time: May 18 10:57:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 06 14:49:39 2020
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1154604	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.17	114	2630777	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.11	117	2516803	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2099240	10.30	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.00%

Target Compounds					Ovalue
29) Chloroform	5.74	83	174757	0.961	ppbv 94
31) cis-1,2-Dichloroethene	5.53	61	438892	4.047	ppbv 97
38) Trichloroethene	7.84	130	6164899	73.368	ppbv 95
52) Tetrachloroethene	11.27	164	12351719	160.153	ppbv 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051420\
Data File : VL035032.D
Acq On : 14 May 2020 22:51
Operator : SY/AP
Sample : L2566-04DL2 1200X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-2DL2

Quant Time: May 18 10:57:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May 06 14:49:39 2020
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

