

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051420\
Data File : VL035036.D
Acq On : 15 May 2020 2:03
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
5/18/2020 1:35:26 PM

Quant Time: May 15 05:58:11 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 0
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1161547	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2638361	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.11	117	2453910	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2047420	10.31	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.10%

Target Compounds					Ovalue
15) Acetone	3.84	43	22429m	0.146	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051420\
Data File : VL035036.D
Acq On : 15 May 2020 2:03
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: May 15 05:58:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
5/18/2020 1:35:26 PM

