


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL051420\
 Data File : VL035017.D
 Acq On : 14 May 2020 11:39
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050620AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed May 06 14:49:39 2020
 Last Update : Wed May 06 14:49:39 2020



AutoFind: Scans 3620, 3621, 3622; Background Corrected with Scan 3597

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	24.9	197930	PASS
75	95	30	66	59.5	472234	PASS
95	95	100	100	100.0	793600	PASS
96	95	5	9	7.2	57245	PASS
173	174	0.00	2	1.0	4852	PASS
174	95	50	120	62.2	493269	PASS
175	174	4	9	7.6	37424	PASS
176	174	93	101	98.4	485290	PASS
177	176	5	9	6.7	32341	PASS