

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051623\
 Data File : VL040435.D
 Acq On : 16 May 2023 12:14
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin
 Yesilyurt

05/17/2023
 Supervised By :Mahesh
 Dadoda

Quant Time: May 16 12:55:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Tue May 16 11:16:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	1009151	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.661	114	2602494	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.481	117	2241967	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.790	95	1793787	9.675	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.919	62	7156	0.104	ppbv	# 65
32) 1,1,1-Trichloroethane	6.005	97	15998m	0.091	ppbv	
35) Carbon Tetrachloride	6.494	117	16673m	0.091	ppbv	
38) Trichloroethene	7.301	130	11032m	0.091	ppbv	
52) Tetrachloroethene	10.635	164	5923m	0.089	ppbv	
54) 1,2-Dibromoethane	10.021	107	7314m	0.079	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.060	83	7685m	0.086	ppbv	
79) Naphthalene	21.429	128	8929m	0.047	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051623\
Data File : VL040435.D
Acq On : 16 May 2023 12:14
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By : Semsettin
Yesilyurt

05/17/2023
Supervised By : Mahesh
Dadoda

05/17/2023

Quant Time: May 16 12:55:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue May 16 11:16:05 2023
Response via : Initial Calibration

