

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051623\
 Data File : VL040446.D
 Acq On : 16 May 2023 20:30
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: May 17 05:17:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed May 17 05:12:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1072482	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.674	114	2871969	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.494	117	2475160	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.806	95	1856544	9.328	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051623\
Data File : VL040446.D
Acq On : 16 May 2023 20:30
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: May 17 05:17:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed May 17 05:12:40 2023
Response via : Initial Calibration

