

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051623\
 Data File : VL040447.D
 Acq On : 16 May 2023 21:12
 Operator : SY/MD
 Sample : 02793-01DL 80X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/17/2023
 Supervised By :Mahesh Dadoda 05/17/2023

Quant Time: May 17 05:17:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed May 17 05:12:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.208	49	1018188	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.671	114	2645454	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.494	117	2279741	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.806	95	1791220	9.772	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.700%
Target Compounds						
					Qvalue	
13) Ethanol	3.266	45	26865m	6.929	ppbv	
15) Acetone	3.494	43	108396	0.842	ppbv	94
20) Methylene Chloride	3.944	84	648942	11.310	ppbv	98
26) Hexane	5.227	57	615001	6.116	ppbv #	35

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL051623\
Data File : VL040447.D
Acq On : 16 May 2023 21:12
Operator : SY/MD
Sample : 02793-01DL 80X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-1DL

Quant Time: May 17 05:17:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL051623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed May 17 05:12:40 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/17/2023
Supervised By : Mahesh Dadoda 05/17/2023

