

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051920\
 Data File : VL035048.D
 Acq On : 19 May 2020 17:40
 Operator : SY/AP
 Sample : L2683-02
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA

Manual Integrations
 APPROVED

MMDadoda
 5/20/2020 4:27:22 PM

Quant Time: May 20 06:03:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 0
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1023504	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2290337	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.10	117	2123749	10.00	ppbv	-0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.43	95	1833480	10.67	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.70%

Target Compounds						Ovalue
2) Dichlorodifluoromethane	3.00	85	32930	0.227	ppbv	97
3) Chlorodifluoromethane	2.94	51	36669m	0.314	ppbv	
4) Chloromethane	3.10	50	52987	0.791	ppbv	89
9) Propene	2.97	41	36947	0.729	ppbv	84
11) Trichlorofluoromethane	3.91	101	30878	0.206	ppbv #	81
13) Ethanol	3.57	45	81256	4.856	ppbv	94
15) Acetone	3.81	43	435979	3.222	ppbv	98
20) Methylene Chloride	4.31	84	25550	0.570	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051920\
Data File : VL035048.D
Acq On : 19 May 2020 17:40
Operator : SY/AP
Sample : L2683-02
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QA

Manual Integrations
APPROVED

MMDadoda
5/20/2020 4:27:22 PM

Quant Time: May 20 06:03:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

