

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051920\
 Data File : VL035071.D
 Acq On : 20 May 2020 8:34
 Operator : SY/AP
 Sample : L2712-02DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

MMDadoda
 5/20/2020 4:28:37 PM

Quant Time: May 20 10:08:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 0
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1201480	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2776403	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.10	117	2621350	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2228754	10.50	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.96	41	11158m	0.187	ppbv	
15) Acetone	3.81	43	1503177	9.463	ppbv	96
17) tert-Butyl alcohol	4.27	59	76007	0.538	ppbv	95
34) 2-Butanone	5.22	43	169425	0.982	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051920\
Data File : VL035071.D
Acq On : 20 May 2020 8:34
Operator : SY/AP
Sample : L2712-02DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Manual Integrations
APPROVED

MMDadoda
5/20/2020 4:28:37 PM

Quant Time: May 20 10:08:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

