


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL051920\
 Data File : VL035041.D
 Acq On : 19 May 2020 11:30
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050620AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed May 06 14:49:39 2020
 Last Update : Wed May 06 14:49:39 2020



AutoFind: Scans 3620, 3621, 3622; Background Corrected with Scan 3597

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	25.5	225429	PASS
75	95	30	66	58.5	517994	PASS
95	95	100	100	100.0	884736	PASS
96	95	5	9	6.9	60997	PASS
173	174	0.00	2	1.0	5393	PASS
174	95	50	120	61.6	545322	PASS
175	174	4	9	7.6	41205	PASS
176	174	93	101	97.3	530624	PASS
177	176	5	9	6.5	34512	PASS