

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052019\
Data File : VL033734.D
Acq On : 20 May 2019 21:36
Operator : SY/AP
Sample : CAN 10044
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10044

Quant Time: May 21 10:03:42 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri May 17 08:15:41 2019
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	855341	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1949960	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2175083	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1795404	10.18	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.80%


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052019\
Data File : VL033734.D
Acq On : 20 May 2019 21:36
Operator : SY/AP
Sample : CAN 10044
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10044

Quant Time: May 21 10:03:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 17 08:15:41 2019
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.76 min Scan# 877

Delta R.T. -0.00 min

Lab File: VL033734.D

Acq: 20 May 2019 21:36

Instrument :

MSVOA_L

ClientSampleId :

CAN 10044


Tot Ion: 49 Resp: 855341

Ion Ratio Lower Upper

49 100

128 50.0 25.2 75.6




130 63.9 32.5 97.5



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.28 min Scan# 1352

Delta R.T. -0.01 min

Lab File: VL033734.D

Acq: 20 May 2019 21:36


Tgt Ion: 114 Resp: 1949960

Ion Ratio Lower Upper

114 100

63 27.1 20.9 31.3



88 17.9 14.3 21.5




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.23 min Scan# 2890
Delta R.T. -0.01 min
Lab File: VL033734.D
Acq: 20 May 2019 21:36

Instrument :
MSVOA_L
ClientSampleId :
CAN 10044

Tot Ion:117 Resp: 2175083
Ion Ratio Lower Upper
117 100
82 67.6 49.4 74.0
119 32.8 24.9 37.3




Abundance Ion 117.10 (116.80 to 117.80): V
Ion 82.10 (81.80 to 82.80): V
Ion 119.10 (118.80 to 119.80): V



#68
1-Bromo-4-Fluorobenzene
Concen: 10.182 ppbv
RT: 14.57 min Scan# 3617
Delta R.T. -0.01 min
Lab File: VL033734.D
Acq: 20 May 2019 21:36

Tgt Ion: 95 Resp: 1795404
Ion Ratio Lower Upper
95 100
174 65.5 53.9 80.9
175 4.6 3.8 5.8



Abundance Ion 95.10 (94.80 to 95.80): V
Ion 174.00 (173.70 to 174.70): V
Ion 175.00 (174.70 to 175.70): V

