


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL052021\
 Data File : VL037018.D
 Acq On : 20 May 2021 8:47
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL051821AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue May 18 12:00:15 2021
 Last Update : Tue May 18 12:00:15 2021



AutoFind: Scans 3603, 3604, 3605; Background Corrected with Scan 3579

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	22.7	164842	PASS
75	95	30	66	55.1	400362	PASS
95	95	100	100	100.0	726293	PASS
96	95	5	9	7.0	50498	PASS
173	174	0.00	2	0.4	1902	PASS
174	95	50	120	70.7	513301	PASS
175	174	4	9	7.3	37328	PASS
176	174	93	101	95.4	489792	PASS
177	176	5	9	7.0	34088	PASS