

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052125\
 Data File : VL042560.D
 Acq On : 21 May 2025 11:27
 Operator : SY/MD
 Sample : MDL 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL 0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/22/2025
 Supervised By :Mahesh Dadoda 05/22/2025

Quant Time: May 21 23:02:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052025AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed May 21 00:21:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.794	49	111905	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.969	114	349175	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.892	117	300601	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.384	95	235594	9.793	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.900%
Target Compounds						
5) Vinyl Chloride	1.583	62	209	0.028	ppbv #	Qvalue 63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052125\
Data File : VL042560.D
Acq On : 21 May 2025 11:27
Operator : SY/MD
Sample : MDL 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Quant Time: May 21 23:02:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052025AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed May 21 00:21:29 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/22/2025
Supervised By : Mahesh Dadoda 05/22/2025

