

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052220\
 Data File : VL035079.D
 Acq On : 22 May 2020 11:23
 Operator : SY/AP
 Sample : VL0522ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0522ABL01

Quant Time: May 22 23:35:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed May 06 14:49:39 2020
 QLast Update : Wed May 06 14:49:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	996207	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	2324402	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.09	117	2242587	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	1889914	10.41	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.83	43	34308	0.260	ppbv	# 65
20) Methylene Chloride	4.31	84	8523m	0.195	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052220\
Data File : VL035079.D
Acq On : 22 May 2020 11:23
Operator : SY/AP
Sample : VL0522ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0522ABL01

Quant Time: May 22 23:35:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL050620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed May 06 14:49:39 2020
QLast Update : Wed May 06 14:49:39 2020
Response via : Initial Calibration

