


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052421\
 Data File : VL037043.D
 Acq On : 24 May 2021 9:33
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL051821AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue May 18 12:00:15 2021
 Last Update : Tue May 18 12:00:15 2021



AutoFind: Scans 3608, 3609, 3610; Background Corrected with Scan 3585

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	21.9	148330	PASS
75	95	30	66	54.4	368298	PASS
95	95	100	100	100.0	677098	PASS
96	95	5	9	6.7	45464	PASS
173	174	0.00	2	0.4	1992	PASS
174	95	50	120	69.2	468864	PASS
175	174	4	9	7.9	37122	PASS
176	174	93	101	97.6	457429	PASS
177	176	5	9	6.3	28882	PASS