

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052424\
 Data File : VL041285.D
 Acq On : 24 May 2024 11:03
 Operator : SY/MD
 Sample : P2599-01 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/28/2024
 Supervised By :Mahesh Dadoda 05/28/2024

Quant Time: May 24 23:19:23 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri May 24 23:11:50 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	1023111	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.664	114	2616690	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2459713	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.780 95 1955560 10.531 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 105.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.754	85	16325	0.118	ppbv	100
3) Chlorodifluoromethane	2.700	51	38067	0.245	ppbv	94
4) Chloromethane	2.835	50	6502	0.107	ppbv	97
9) Propene	2.719	41	209610	3.447	ppbv	93
10) Heptane	7.584	43	56830m	0.386	ppbv	
11) Trichlorofluoromethane	3.581	101	3603787	25.622	ppbv	98
13) Ethanol	3.266	45	56435m	4.140	ppbv	
15) Acetone	3.484	43	2168777	17.961	ppbv	92
17) tert-Butyl alcohol	3.902	59	28365	0.218	ppbv #	81
20) Methylene Chloride	3.951	84	96135	2.290	ppbv	92
26) Hexane	5.224	57	271247	2.522	ppbv #	39
27) Carbon Disulfide	4.143	76	174387	1.370	ppbv	99
29) Chloroform	5.288	83	28737m	0.194	ppbv	
30) Cyclohexane	6.613	84	28465	0.334	ppbv #	74
32) 1,1,1-Trichloroethane	6.012	97	13576m	0.092	ppbv	
34) 2-Butanone	4.799	43	2300575	17.274	ppbv	99
36) Benzene	6.378	78	55222	0.251	ppbv #	89
38) Trichloroethene	7.311	130	4280m	0.052	ppbv	
44) 2,2,4-Trimethylpentane	7.333	57	116684	0.308	ppbv #	72
51) 2-Hexanone	9.564	43	218496	1.232	ppbv #	96
52) Tetrachloroethene	10.622	164	349117	4.523	ppbv	94
53) Toluene	9.230	91	294839	1.158	ppbv	99
58) Ethyl Benzene	12.095	91	63173m	0.187	ppbv	
59) m/p-Xylene	12.349	91	170687m	0.602	ppbv	
60) o-Xylene	13.056	91	102245	0.385	ppbv	100
65) tert-Butylbenzene	16.111	119	34050m	0.096	ppbv	
72) 4-Ethyltoluene	15.194	105	58712m	0.195	ppbv	
73) 1,3,5-Trimethylbenzene	15.355	105	91426	0.336	ppbv	99
74) 1,2,4-Trimethylbenzene	16.108	105	290684	0.977	ppbv #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052424\
Data File : VL041285.D
Acq On : 24 May 2024 11:03
Operator : SY/MD
Sample : P2599-01 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
SV1

Quant Time: May 24 23:19:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 24 23:11:50 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By : Semsettin Yesilyurt 05/28/2024
Supervised By : Mahesh Dadoda 05/28/2024





#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 5.205 min Scan# 804
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 49 Resp: 102311
Ion Ratio Lower Upper
49 100
128 45.5 21.5 64.5
130 59.3 27.2 81.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#2
Dichlorodifluoromethane
Concen: 0.118 ppbv
RT: 2.754 min Scan# 47
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 85 Resp: 16325
Ion Ratio Lower Upper
85 100
87 32.0 25.7 38.5





#3
Chlorodifluoromethane
Concen: 0.245 ppbv
RT: 2.700 min Scan# 30
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 51 Resp: 3806
Ion Ratio Lower Upper
51 100
67 14.9 14.0 21.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#4
Chloromethane
Concen: 0.107 ppbv
RT: 2.835 min Scan# 72
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 50 Resp: 6502
Ion Ratio Lower Upper
50 100
52 37.3 28.3 42.5





#9
 Propene
 Concen: 3.447 ppbv
 RT: 2.719 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: VL041285.D
 Acq: 24 May 2024 11:03




Instrument :
 MSVOA_L
 ClientSampleId :
 SV1



Tgt Ion: 41 Resp: 209610
 Ion Ratio Lower Upper
 41 100
 42 63.5 55.3 82.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
 Supervised By :Mahesh Dadoda 05/28/2024



#10
 Heptane
 Concen: 0.386 ppbv m
 RT: 7.584 min Scan# 1549
 Delta R.T. -0.006 min
 Lab File: VL041285.D
 Acq: 24 May 2024 11:03


Tgt Ion: 43 Resp: 56830
 Ion Ratio Lower Upper
 43 100
 57 0.0 38.7 58.1#





#11
Trichlorofluoromethane
Concen: 25.622 ppbv
RT: 3.581 min Scan# 304
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion:101 Resp: 360378
Ion Ratio Lower Upper
101 100
103 66.2 51.6 77.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#13
Ethanol
Concen: 4.140 ppbv m
RT: 3.266 min Scan# 206
Delta R.T. 0.006 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 45 Resp: 56435
Ion Ratio Lower Upper
45 100
46 38.8 14.8 59.0





#15
Acetone
Concen: 17.961 ppbv
RT: 3.484 min Scan# 21
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 43 Resp: 216877
Ion Ratio Lower Upper
43 100
58 30.9 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#17
tert-Butyl alcohol
Concen: 0.218 ppbv
RT: 3.902 min Scan# 404
Delta R.T. 0.006 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 59 Resp: 28365
Ion Ratio Lower Upper
59 100
57 17.7 8.4 12.6#





#20
Methylene Chloride
Concen: 2.290 ppbv
RT: 3.951 min Scan# 411
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 84 Resp: 96135
Ion Ratio Lower Upper
84 100
49 149.8 130.4 195.6
51 47.7 38.0 57.0
86 54.5 50.3 75.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#26
Hexane
Concen: 2.522 ppbv
RT: 5.224 min Scan# 815
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 57 Resp: 271247
Ion Ratio Lower Upper
57 100
41 100.8 75.8 113.8
43 73.1 187.0 280.4#
56 58.3 42.7 64.1





#27
Carbon Disulfide
Concen: 1.370 ppbv
RT: 4.143 min Scan# 41
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 76 Resp: 17438
Ion Ratio Lower Upper
76 100
44 20.8 16.3 24.5
78 9.9 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#29
Chloroform
Concen: 0.194 ppbv m
RT: 5.288 min Scan# 835
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Tgt Ion: 83 Resp: 28737
Ion Ratio Lower Upper
83 100
85 67.5 51.7 77.5









#30
Cyclohexane
Concen: 0.334 ppbv
RT: 6.613 min Scan# 1060
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Instrument :
MSVOA_L
ClientSampleId :
SV1

Tgt Ion: 84 Resp: 2846
Ion Ratio Lower Upper
84 100
56 132.3 116.8 175.2
41 140.7 73.8 110.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#32
1,1,1-Trichloroethane
Concen: 0.092 ppbv m
RT: 6.012 min Scan# 1060
Delta R.T. -0.006 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Tgt Ion: 97 Resp: 13576
Ion Ratio Lower Upper
97 100
99 65.8 51.7 77.5
61 50.8 41.1 61.7





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 6.664 min Scan# 11

Delta R.T. -0.003 min

Lab File: VL041285.D

Acq: 24 May 2024 11:03

Instrument :

MSVOA_L

ClientSampleId :

SV1

Tgt Ion:114 Resp: 2616690

Ion Ratio Lower Upper

114 100

63 26.0 21.8 32.6





88 18.0 15.2 22.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 05/28/2024

Supervised By :Mahesh Dadoda 05/28/2024



#34

2-Butanone

Concen: 17.274 ppbv

RT: 4.799 min Scan# 683

Delta R.T. -0.003 min

Lab File: VL041285.D




Acq: 24 May 2024 11:03


Tgt Ion: 43 Resp: 2300575

Ion Ratio Lower Upper

43 100

72 22.2 17.3 25.9





#36

Benzene

Concen: 0.251 ppbv

RT: 6.378 min Scan# 1171

Delta R.T. -0.006 min

Lab File: VL041285.D

Acq: 24 May 2024 11:03

Instrument :

MSVOA_L

ClientSampleId :





SV1

Tgt Ion: 78	Resp: 5522
Ion Ratio	Lower Upper
78 100	
77 17.0	18.3 27.5
50 15.7	15.9 23.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#38

Trichloroethene

Concen: 0.052 ppbv m




RT: 7.311 min Scan# 1464


Delta R.T. 0.010 min

Lab File: VL041285.D

Acq: 24 May 2024 11:03


Tgt Ion:130	Resp: 4280
Ion Ratio	Lower Upper
130 100	
95 135.6	95.0 142.6
132 145.2	77.0 115.4
97 59.8	59.1 88.7





#44
2,2,4-Trimethylpentane
Concen: 0.308 ppbv
RT: 7.333 min Scan# 1468
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 57 Resp: 116684
Ion Ratio Lower Upper
57 100
41 51.5 29.5 44.3
56 49.6 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#51
2-Hexanone
Concen: 1.232 ppbv
RT: 9.564 min Scan# 2165
Delta R.T. 0.006 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 43 Resp: 218496
Ion Ratio Lower Upper
43 100
58 45.4 38.4 57.6
98 0.3 0.0 0.0





#52
Tetrachloroethene
Concen: 4.523 ppbv
RT: 10.622 min Scan# 2494
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion:164 Resp: 34911
Ion Ratio Lower Upper
164 100
166 124.3 101.1 151.7
129 114.6 81.7 122.5
131 104.9 80.4 120.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#53
Toluene
Concen: 1.158 ppbv
RT: 9.230 min Scan# 2061
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Tgt Ion: 91 Resp: 294839
Ion Ratio Lower Upper
91 100
92 61.0 48.2 72.4









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 11.471 min Scan# 21
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Instrument :
MSVOA_L
ClientSampleId :
SV1

Tgt Ion:117 Resp: 245971
Ion Ratio Lower Upper
117 100
82 64.7 53.6 80.4
119 32.8 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#58
Ethyl Benzene
Concen: 0.187 ppbv m
RT: 12.095 min Scan# 2952
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Tgt Ion: 91 Resp: 63173
Ion Ratio Lower Upper
91 100
106 33.7 24.3 36.5









#59
m/p-Xylene
Concen: 0.602 ppbv m
RT: 12.349 min Scan# 3031
Delta R.T. -0.029 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Instrument :
MSVOA_L
ClientSampleId :
SV1

Tgt Ion: 91 Resp: 17068
Ion Ratio Lower Upper
91 100
106 0.0 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#60
o-Xylene
Concen: 0.385 ppbv
RT: 13.056 min Scan# 3251
Delta R.T. -0.006 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion: 91 Resp: 102245
Ion Ratio Lower Upper
91 100
106 45.0 22.4 67.3





#65
tert-Butylbenzene
Concen: 0.096 ppbv
RT: 16.111 min Scan# 4195
Delta R.T. 0.019 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03




Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion: 119 Resp: 34050
Ion Ratio Lower Upper
119 100
91 116.3 67.0 100.4
134 0.0 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#68
1-Bromo-4-Fluorobenzene
Concen: 10.531 ppbv
RT: 13.780 min Scan# 3476
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Tgt Ion: 95 Resp: 1955560
Ion Ratio Lower Upper
95 100
174 66.2 53.6 80.4
175 4.9 4.2 6.4









#72
4-Ethyltoluene
Concen: 0.195 ppbv m
RT: 15.194 min Scan# 3915
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Instrument :
MSVOA_L
ClientSampleId :
SV1

Tgt Ion:105 Resp: 58712
Ion Ratio Lower Upper
105 100
120 32.2 24.7 37.1
91 14.5 10.3 15.5
77 21.2 11.4 17.0#




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024



#73
1,3,5-Trimethylbenzene
Concen: 0.336 ppbv
RT: 15.355 min Scan# 3966
Delta R.T. -0.000 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03


Tgt Ion:105 Resp: 91426
Ion Ratio Lower Upper
105 100
120 45.6 37.0 55.4





#74
1,2,4-Trimethylbenzene
Concen: 0.977 ppbv
RT: 16.108 min Scan# 41
Delta R.T. -0.003 min
Lab File: VL041285.D
Acq: 24 May 2024 11:03

Instrument :
MSVOA_L
ClientSampleId :
SV1



Tgt Ion:	105	Resp:	290684
Ion	Ratio	Lower	Upper
105	100		
120	40.4	37.8	56.6
91	10.6	44.3	66.5
77	12.9	19.0	28.4

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/28/2024
Supervised By :Mahesh Dadoda 05/28/2024

