

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052522\
 Data File : VL039130.D
 Acq On : 26 May 2022 11:18
 Operator : SY/AP
 Sample : QC052522 CAN 10294
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC052522 CAN 10294

Quant Time: May 27 04:28:18 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050522AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 05 10:12:15 2022
 QLast Update : Thu May 05 10:12:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.645	49	909952	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	7.147	114	2075307	10.000	ppbv	-0.03
55) Chlorobenzene-d5	12.050	117	1685216	10.000	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.378	95	1482167	11.106	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	111.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052522\
Data File : VL039130.D
Acq On : 26 May 2022 11:18
Operator : SY/AP
Sample : QC052522 CAN 10294
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC052522 CAN 10294

Quant Time: May 27 04:28:18 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL050522AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu May 05 10:12:15 2022
QLast Update : Thu May 05 10:12:15 2022
Response via : Initial Calibration

