

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052819\
 Data File : VL033824.D
 Acq On : 29 May 2019 5:05
 Operator : SY/AP
 Sample : K2960-06 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BPS1-SG-3009-25

Manual Integrations
 APPROVED

apatel
 5/29/2019 12:41:25 PM

Quant Time: May 29 07:03:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 1
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	823518	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1955976	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1754965	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1460902	10.27	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.70%

Target Compounds					Ovalue
9) Propene	3.05	41	11848	0.295 ppbv	88
13) Ethanol	3.66	45	5819m	0.449 ppbv	
15) Acetone	3.90	43	73741	0.645 ppbv	96
20) Methylene Chloride	4.41	84	7723m	0.176 ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052819\
Data File : VL033824.D
Acq On : 29 May 2019 5:05
Operator : SY/AP
Sample : K2960-06 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BPS1-SG-3009-25

Manual Integrations
APPROVED

apatel
5/29/2019 12:41:25 PM

Quant Time: May 29 07:03:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration

