

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052819\
 Data File : VL033827.D
 Acq On : 29 May 2019 7:00
 Operator : SY/AP
 Sample : K2960-08DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BPS1-SG-3010-55DL

Manual Integrations
 APPROVED

apatel
 5/29/2019 12:41:35 PM

Quant Time: May 29 08:18:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 1
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	819662	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1919491	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1740822	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1463588	10.37	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	64324	1.611	ppbv	91
11) Trichlorofluoromethane	4.00	101	20071	0.122	ppbv	90
13) Ethanol	3.66	45	10539m	0.817	ppbv	
15) Acetone	3.90	43	454693	3.998	ppbv	100
34) 2-Butanone	5.35	43	22970m	0.193	ppbv	
53) Toluene	9.94	91	22208	0.118	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052819\
Data File : VL033827.D
Acq On : 29 May 2019 7:00
Operator : SY/AP
Sample : K2960-08DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BPS1-SG-3010-55DL

Manual Integrations
APPROVED

apatel
5/29/2019 12:41:35 PM

Quant Time: May 29 08:18:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration

