

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052819\
 Data File : VL033829.D
 Acq On : 29 May 2019 8:17
 Operator : SY/AP
 Sample : K2960-01DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BPS1-SG-3010-25DL

Manual Integrations
 APPROVED

apatel
 5/29/2019 12:41:42 PM

Quant Time: May 29 09:00:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 1
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	817679	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1873906	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1762047	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1423745	9.97	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	30609	0.768	ppbv	92
13) Ethanol	3.67	45	13063m	1.015	ppbv	
15) Acetone	3.90	43	437725	3.858	ppbv	99
53) Toluene	9.94	91	19114m	0.104	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052819\
Data File : VL033829.D
Acq On : 29 May 2019 8:17
Operator : SY/AP
Sample : K2960-01DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
BPS1-SG-3010-25DL

Manual Integrations
APPROVED

apatel
5/29/2019 12:41:42 PM

Quant Time: May 29 09:00:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration

