


Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052819\  
 Data File : VL033810.D  
 Acq On : 28 May 2019 18:34  
 Operator : SY/AP  
 Sample : BFB  
 Misc : 400ml/MSVOA L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_L\METHODS\VL051719AIR.M  
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA\_L Fri May 17 08:15:41 2019  
 Last Update : Fri May 17 08:15:41 2019



AutoFind: Scans 3616, 3617, 3618; Background Corrected with Scan 3588

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 8            | 40           | 21.6      | 168106  | PASS             |
| 75          | 95           | 30           | 66           | 53.9      | 419434  | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 777557  | PASS             |
| 96          | 95           | 5            | 9            | 6.5       | 50637   | PASS             |
| 173         | 174          | 0.00         | 2            | 0.8       | 4391    | PASS             |
| 174         | 95           | 50           | 120          | 68.5      | 532352  | PASS             |
| 175         | 174          | 4            | 9            | 7.2       | 38376   | PASS             |
| 176         | 174          | 93           | 101          | 97.5      | 518933  | PASS             |
| 177         | 176          | 5            | 9            | 6.4       | 33389   | PASS             |