

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060122\
 Data File : VL039151.D
 Acq On : 1 Jun 2022 17:09
 Operator : SY/AP
 Sample : N3106-01 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/06/2022
 Supervised By :Mahesh Dadoda 06/06/2022

Quant Time: Jun 02 00:40:29 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Jun 01 01:22:18 2022

QLast Update : Wed Jun 01 01:22:18 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	1240372	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.179	114	3256702	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.089	117	2999023	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.417	95	2566770	10.352	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.500%
Target Compounds						
					Qvalue	
9) Propene	3.031	41	4750513m	69.031	ppbv	
10) Heptane	8.131	43	5161300	32.150	ppbv	98
13) Ethanol	3.616	45	12933m	3.446	ppbv	
20) Methylene Chloride	4.353	84	137451	2.524	ppbv	91
26) Hexane	5.697	57	8687158	67.881	ppbv #	39
34) 2-Butanone	5.256	43	230375m	1.302	ppbv	
52) Tetrachloroethene	11.230	164	236644m	2.292	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060122\
Data File : VL039151.D
Acq On : 1 Jun 2022 17:09
Operator : SY/AP
Sample : N3106-01 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1

Quant Time: Jun 02 00:40:29 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/06/2022
Supervised By :Mahesh Dadoda 06/06/2022

