

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060221\
 Data File : VL037105.D
 Acq On : 2 Jun 2021 17:12
 Operator : SY/AP
 Sample : M2548-09DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-9ADL

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 3:00:39 PM

Quant Time: Jun 06 20:08:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1403743	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4125137	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3706306	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2659599	9.55	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	17694	0.192	ppbv	96
13) Ethanol	3.59	45	2701303	219.855	ppbv	92
15) Acetone	3.85	43	175873m	1.223	ppbv	
26) Hexane	5.67	57	86114m	0.488	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060221\
Data File : VL037105.D
Acq On : 2 Jun 2021 17:12
Operator : SY/AP
Sample : M2548-09DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-9ADL

Manual Integrations
APPROVED

MMDadoda
6/7/2021 3:00:39 PM

Quant Time: Jun 06 20:08:47 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

