

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060221\
 Data File : VL037113.D
 Acq On : 2 Jun 2021 22:53
 Operator : SY/AP
 Sample : M2548-13DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SGS-2ADL

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 3:01:16 PM

Quant Time: Jun 06 20:22:41 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1358072	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4020512	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3608405	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2664752	9.83	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	64037m	5.387	ppbv	
15) Acetone	3.84	43	283664	2.039	ppbv	# 82
20) Methylene Chloride	4.33	84	265417	3.560	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060221\
Data File : VL037113.D
Acq On : 2 Jun 2021 22:53
Operator : SY/AP
Sample : M2548-13DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SGS-2ADL

Manual Integrations
APPROVED

MMDadoda
6/7/2021 3:01:16 PM

Quant Time: Jun 06 20:22:41 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

