

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060221\
 Data File : VL037115.D
 Acq On : 3 Jun 2021 00:12
 Operator : SY/AP
 Sample : M2548-14DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SGS-5ADL

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 3:01:20 PM

Quant Time: Jun 06 20:26:06 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1338654	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4104557	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3632885	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2673718	9.80	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	205690	17.555	ppbv	96
15) Acetone	3.84	43	381210	2.780	ppbv	90
19) Isopropyl Alcohol	3.99	45	20516m	0.278	ppbv	
20) Methylene Chloride	4.33	84	173707	2.363	ppbv	98
26) Hexane	5.67	57	116251	0.691	ppbv #	46

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060221\
Data File : VL037115.D
Acq On : 3 Jun 2021 00:12
Operator : SY/AP
Sample : M2548-14DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SGS-5ADL

Manual Integrations
APPROVED

MMDadoda
6/7/2021 3:01:20 PM

Quant Time: Jun 06 20:26:06 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

