

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060221\
 Data File : VL037117.D
 Acq On : 3 Jun 2021 1:31
 Operator : SY/AP
 Sample : M2548-15DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SGS-6ADL

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 3:01:24 PM

Quant Time: Jun 06 20:30:31 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1394862	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4067216	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3649850	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2660277	9.70	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	33709m	2.761	ppbv	
15) Acetone	3.86	43	173193	1.212	ppbv #	77
19) Isopropyl Alcohol	3.99	45	117408	1.529	ppbv #	43
34) 2-Butanone	5.27	43	24297m	0.164	ppbv	
38) Trichloroethene	7.82	130	38436m	0.256	ppbv	
52) Tetrachloroethene	11.20	164	4517m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060221\
Data File : VL037117.D
Acq On : 3 Jun 2021 1:31
Operator : SY/AP
Sample : M2548-15DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SGS-6ADL

Manual Integrations
APPROVED

MMDadoda
6/7/2021 3:01:24 PM

Quant Time: Jun 06 20:30:31 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

