

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060322\
 Data File : VL039158.D
 Acq On : 3 Jun 2022 18:06
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/05/2022
 Supervised By : Mahesh Dadoda 06/06/2022

Quant Time: Jun 04 00:17:56 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 01 01:22:18 2022

QLast Update : Wed Jun 01 01:22:18 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.681	49	1294601	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.189	114	3639624	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.098	117	3118185	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.426	95	2310630	8.963	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	89.600%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	3.272	62	6558m	0.083	ppbv	
9) Propene	3.025	41	16964m	0.236	ppbv	
32) 1,1,1-Trichloroethane	6.533	97	13946m	0.063	ppbv	
35) Carbon Tetrachloride	7.038	117	14299m	0.054	ppbv	
38) Trichloroethene	7.848	130	11726m	0.086	ppbv	
52) Tetrachloroethene	11.233	164	8259m	0.072	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.693	83	12897m	0.065	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060322\
Data File : VL039158.D
Acq On : 3 Jun 2022 18:06
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 06/05/2022
Supervised By : Mahesh Dadoda 06/06/2022

Quant Time: Jun 04 00:17:56 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

