

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060424\
 Data File : VL041298.D
 Acq On : 4 Jun 2024 9:11
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024
 Supervised By :Mahesh Dadoda 06/05/2024

Quant Time: Jun 04 11:29:47 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL060424AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri May 24 23:11:50 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	1023215	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	2669645	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2568407m	10.000	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.770 95 1978642 10.204 ppbv -0.02
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.000%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.758	85	1473180	10.644	ppbv	98
3) Chlorodifluoromethane	2.703	51	1456695	9.374	ppbv	98
4) Chloromethane	2.838	50	554948	9.101	ppbv	94
5) Vinyl Chloride	2.944	62	570339	9.310	ppbv	98
6) Bromomethane	3.144	94	196997	8.814	ppbv	89
7) Chloroethane	3.218	64	198033	9.156	ppbv	94
8) Dichlorotetrafluoroethane	2.883	85	1373725	9.589	ppbv	100
9) Propene	2.722	41	516622	8.495	ppbv	100
10) Heptane	7.581	43	1292998	8.775	ppbv	100
11) Trichlorofluoromethane	3.584	101	1382785	9.830	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.082	101	992255	9.539	ppbv	97
13) Ethanol	3.263	45	126601	9.286	ppbv	97
14) Bromoethene	3.388	108	413913	9.642	ppbv	97
15) Acetone	3.488	43	1039645	8.609	ppbv	98
16) 1,3-Butadiene	3.009	39	580608	9.525	ppbv	98
17) tert-Butyl alcohol	3.893	59	1072137	8.230	ppbv	98
18) 1,1-Dichloroethene	3.896	96	439214	9.470	ppbv	95
19) Isopropyl Alcohol	3.594	45	764874	9.653	ppbv #	44
20) Methylene Chloride	3.951	84	357513	8.514	ppbv	95
21) Allyl Chloride	4.012	41	706520	8.973	ppbv	97
22) trans-1,2-Dichloroethene	4.455	96	455220	9.438	ppbv	97
23) Vinyl Acetate	4.639	43	1227126	9.228	ppbv	99
24) 1,1-Dichloroethane	4.578	63	982654	9.504	ppbv	99
25) Ethyl Acetate	5.214	43	2140766	8.882	ppbv	99
26) Hexane	5.221	57	930822	8.654	ppbv	99
27) Carbon Disulfide	4.140	76	1158452	9.098	ppbv	97
28) Methyl tert-Butyl Ether	4.597	73	720670	8.414	ppbv	99
29) Chloroform	5.285	83	1465284	9.875	ppbv	97
30) Cyclohexane	6.610	84	785466	9.225	ppbv	99
31) cis-1,2-Dichloroethene	5.089	61	944077	9.375	ppbv	95
32) 1,1,1-Trichloroethane	6.012	97	1443296	9.748	ppbv	98
34) 2-Butanone	4.796	43	1247580	9.182	ppbv	100
35) Carbon Tetrachloride	6.500	117	1496089	9.986	ppbv	99
36) Benzene	6.378	78	1980600	8.841	ppbv	99
37) 1,2-Dichloroethane	5.816	62	1159815	9.922	ppbv	99
38) Trichloroethene	7.295	130	752906	9.018	ppbv	99
39) 1,2-Dichloropropane	7.079	63	766472	8.877	ppbv	97
40) 1,4-Dioxane	7.295	88	285688	9.262	ppbv	100
41) Tetrahydrofuran	5.565	42	814407	8.279	ppbv	98
42) Bromodichloromethane	7.256	83	1685711	10.055	ppbv	100
43) Methyl Methacrylate	7.478	69	810686	8.947	ppbv	99
44) 2,2,4-Trimethylpentane	7.333	57	3339096	8.644	ppbv	100
45) t-1,3-Dichloropropene	8.713	75	598199	9.110	ppbv	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060424\
 Data File : VL041298.D
 Acq On : 4 Jun 2024 9:11
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/05/2024
 Supervised By : Mahesh Dadoda 06/05/2024

Quant Time: Jun 04 11:29:47 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL060424AIR.M


Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri May 24 23:11:50 2024

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.143	75	709965	8.832	ppbv	99
47) 1,1,2-Trichloroethane	8.899	97	825142	9.548	ppbv	96
48) Dibromochloromethane	9.709	129	1404173	10.198	ppbv	100
49) Bromoform	12.404	173	1158551	9.891	ppbv	99
50) 4-Methyl-2-Pentanone	8.179	43	2102376	8.622	ppbv	100
51) 2-Hexanone	9.549	43	1551244	8.576	ppbv #	98
52) Tetrachloroethene	10.613	164	707765	8.988	ppbv	97
53) Toluene	9.221	91	2322300	8.940	ppbv	98
54) 1,2-Dibromoethane	10.008	107	1126770	9.626	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.507	131	1026568	9.857	ppbv	99
57) Chlorobenzene	11.526	112	1746611	9.286	ppbv	96
58) Ethyl Benzene	12.082	91	3126554	8.878	ppbv	97
59) m/p-Xylene	12.365	91	5307476m	17.931	ppbv	
60) o-Xylene	13.053	91	2493688	8.988	ppbv	99
61) Styrene	12.892	104	1348748	9.202	ppbv	99
62) Isopropylbenzene	14.021	105	3738569	9.008	ppbv	99
63) 1,1,2,2-Tetrachloroethane	13.044	83	1786757	9.448	ppbv	100
64) n-propylbenzene	14.908	120	995001	9.610	ppbv	95
65) tert-Butylbenzene	16.082	119	3310807	8.907	ppbv	99
66) Benzyl Chloride	16.323	91	270848m	6.437	ppbv	
67) sec-Butylbenzene	16.629	105	4696700	8.976	ppbv	99
69) p-Isopropyltoluene	16.989	119	3901704	9.048	ppbv	100
70) n-Butylbenzene	17.876	91	4087151	9.076	ppbv	99
71) 2-Chlorotoluene	14.796	91	2905543	9.195	ppbv	99
72) 4-Ethyltoluene	15.188	105	3019826	9.609	ppbv	98
73) 1,3,5-Trimethylbenzene	15.342	105	2612307	9.187	ppbv	98
74) 1,2,4-Trimethylbenzene	16.105	105	2859255	9.207	ppbv	92
75) 1,3-Dichlorobenzene	16.323	146	1722261	9.672	ppbv	99
76) 1,4-Dichlorobenzene	16.465	146	1671398	9.760	ppbv	99
77) 1,2-Dichlorobenzene	17.137	146	1666911	9.365	ppbv	99
78) Hexachloro-1,3-Butadiene	22.625	225	1207700	7.634	ppbv	99
79) Naphthalene	21.371	128	3074007	10.395	ppbv	99
80) Naphthalene,2-methyl-	24.365	142	1054655	8.865	ppbv	99
81) 1,2,4-Trichlorobenzene	21.130	180	1450918m	9.292	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 5.205 min Scan# 809
Delta R.T. -0.000 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 49 Resp: 1023215
Ion Ratio Lower Upper
49 100
128 45.5 21.5 64.5
130 56.9 27.2 81.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#2
Dichlorodifluoromethane
Concen: 10.644 ppbv
RT: 2.758 min Scan# 48
Delta R.T. 0.013 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11



Tgt Ion: 85 Resp: 1473180
Ion Ratio Lower Upper
85 100
87 33.1 25.7 38.5









#3
Chlorodifluoromethane
Concen: 9.374 ppbv
RT: 2.703 min Scan# 31
Delta R.T. 0.013 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 51 Resp: 145669
Ion Ratio Lower Upper
51 100
67 18.4 14.0 21.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#4
Chloromethane
Concen: 9.101 ppbv
RT: 2.838 min Scan# 73
Delta R.T. 0.010 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 50 Resp: 554948
Ion Ratio Lower Upper
50 100
52 32.0 28.3 42.5





#5
 Vinyl Chloride
 Concen: 9.310 ppbv
 RT: 2.944 min Scan# 106
 Delta R.T. 0.010 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010




Tgt Ion: 62 Resp: 570339
 Ion Ratio Lower Upper
 62 100
 64 33.3 25.7 38.5

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
 Supervised By :Mahesh Dadoda 06/05/2024




#6
 Bromomethane
 Concen: 8.814 ppbv
 RT: 3.144 min Scan# 168
 Delta R.T. 0.010 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11




Tgt Ion: 94 Resp: 196997
 Ion Ratio Lower Upper
 94 100
 96 98.2 70.1 105.1





#7
Chloroethane
Concen: 9.156 ppbv
RT: 3.218 min Scan# 191
Delta R.T. 0.003 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument : MSVOA_L
ClientSampleId : VSTDICCC010



Tgt Ion: 64 Resp: 198033
Ion Ratio Lower Upper
64 100
66 35.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#8
Dichlorotetrafluoroethane
Concen: 9.589 ppbv
RT: 2.883 min Scan# 87
Delta R.T. 0.010 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 85 Resp: 1373725
Ion Ratio Lower Upper
85 100
135 64.6 51.5 77.3





#9
 Propene
 Concen: 8.495 ppbv
 RT: 2.722 min Scan# 31
 Delta R.T. 0.013 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010




Tgt Ion: 41 Resp: 51662
 Ion Ratio Lower Upper
 41 100
 42 69.0 55.3 82.9

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
 Supervised By :Mahesh Dadoda 06/05/2024




#10
 Heptane
 Concen: 8.775 ppbv
 RT: 7.581 min Scan# 1548
 Delta R.T. -0.016 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11




Tgt Ion: 43 Resp: 1292998
 Ion Ratio Lower Upper
 43 100
 57 48.7 38.7 58.1





#11
Trichlorofluoromethane
Concen: 9.830 ppbv
RT: 3.584 min Scan# 305
Delta R.T. 0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion:101 Resp: 138278
Ion Ratio Lower Upper
101 100
103 63.3 51.6 77.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#12
1,1,2-Trichlorotrifluoroethane
Concen: 9.539 ppbv
RT: 4.082 min Scan# 460
Delta R.T. 0.003 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion:101 Resp: 992255
Ion Ratio Lower Upper
101 100
151 68.5 56.8 85.2





#13
Ethanol
Concen: 9.286 ppbv
RT: 3.263 min Scan# 205
Delta R.T. 0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 45 Resp: 12660
Ion Ratio Lower Upper
45 100
46 35.4 14.8 59.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#14
Bromoethene
Concen: 9.642 ppbv
RT: 3.388 min Scan# 244
Delta R.T. 0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion:108 Resp: 413913
Ion Ratio Lower Upper
108 100
106 106.7 88.2 132.4
79 22.3 17.4 26.0





#15
Acetone
Concen: 8.609 ppbv
RT: 3.488 min Scan# 21
Delta R.T. 0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 43 Resp: 103964
Ion Ratio Lower Upper
43 100
58 25.8 21.4 32.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#16
1,3-Butadiene
Concen: 9.525 ppbv
RT: 3.009 min Scan# 126
Delta R.T. 0.010 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion: 39 Resp: 580608
Ion Ratio Lower Upper
39 100
54 75.6 62.2 93.4





#17
tert-Butyl alcohol
Concen: 8.230 ppbv
RT: 3.893 min Scan# 401
Delta R.T. 0.003 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 59 Resp: 107213
Ion Ratio Lower Upper
59 100
57 11.2 8.4 12.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#18
1,1-Dichloroethene
Concen: 9.470 ppbv
RT: 3.896 min Scan# 402
Delta R.T. 0.000 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion: 96 Resp: 439214
Ion Ratio Lower Upper
96 100
61 214.8 164.4 246.6
98 66.4 51.8 77.8





#19
Isopropyl Alcohol
Concen: 9.653 ppbv
RT: 3.594 min Scan# 308
Delta R.T. 0.003 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 45 Resp: 764874
Ion Ratio Lower Upper
45 100
58 1.8 26.8 40.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#20
Methylene Chloride
Concen: 8.514 ppbv
RT: 3.951 min Scan# 419
Delta R.T. 0.003 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11



Tgt Ion: 84 Resp: 357513
Ion Ratio Lower Upper
84 100
49 170.2 130.4 195.6
51 52.4 38.0 57.0
86 65.7 50.3 75.5





#21

Allyl Chloride

Concen: 8.973 ppbv

RT: 4.012 min Scan# 41

Delta R.T. 0.000 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion: 41 Resp: 706520

Ion Ratio Lower Upper

41 100

76 27.9 23.7 35.5





39 82.9 64.7 97.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#22

trans-1,2-Dichloroethene

Concen: 9.438 ppbv

RT: 4.455 min Scan# 576

Delta R.T. -0.006 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11




Tgt Ion: 96 Resp: 455220


Ion Ratio Lower Upper

96 100

61 169.4 137.6 206.4

98 59.5 50.6 76.0









#23
 Vinyl Acetate
 Concen: 9.228 ppbv
 RT: 4.639 min Scan# 613
 Delta R.T. 0.000 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Tgt Ion: 43 Resp: 1227120
 Ion Ratio Lower Upper
 43 100
 86 5.9 4.9 7.3
 42 10.0 8.0 12.0
 44 5.3 3.7 5.5




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
 Supervised By :Mahesh Dadoda 06/05/2024



#24
 1,1-Dichloroethane
 Concen: 9.504 ppbv
 RT: 4.578 min Scan# 614
 Delta R.T. 0.000 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11


Tgt Ion: 63 Resp: 982654
 Ion Ratio Lower Upper
 63 100
 98 3.8 1.6 4.8
 100 2.3 1.1 3.5





#25
Ethyl Acetate
Concen: 8.882 ppbv
RT: 5.214 min Scan# 812
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion: 43 Resp: 2140760
Ion Ratio Lower Upper
43 100
45 10.1 8.2 12.2
61 9.2 7.7 11.5
70 6.9 5.8 8.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#26
Hexane
Concen: 8.654 ppbv
RT: 5.221 min Scan# 814
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 57 Resp: 930822
Ion Ratio Lower Upper
57 100
41 96.4 75.8 113.8
43 232.6 187.0 280.4
56 53.6 42.7 64.1





#27
Carbon Disulfide
Concen: 9.098 ppbv
RT: 4.140 min Scan# 41
Delta R.T. 0.000 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion: 76 Resp: 115845
Ion Ratio Lower Upper
76 100
44 22.1 16.3 24.5
78 9.7 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#28
Methyl tert-Butyl Ether
Concen: 8.414 ppbv
RT: 4.597 min Scan# 620
Delta R.T. -0.003 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Tgt Ion: 73 Resp: 720670
Ion Ratio Lower Upper
73 100
57 24.0 18.8 28.2





#29

Chloroform

Concen: 9.875 ppbv

RT: 5.285 min Scan# 834

Delta R.T. -0.003 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion: 83 Resp: 1465284

Ion Ratio Lower Upper

83 100





85 66.6 51.7 77.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#30

Cyclohexane

Concen: 9.225 ppbv

RT: 6.610 min Scan# 1246

Delta R.T. -0.006 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11




Tgt Ion: 84 Resp: 785466


Ion Ratio Lower Upper

84 100

56 143.6 116.8 175.2

41 93.0 73.8 110.8





#31

cis-1,2-Dichloroethene

Concen: 9.375 ppbv

RT: 5.089 min Scan# 71

Delta R.T. -0.006 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion: 61 Resp: 94407

Ion Ratio Lower Upper

61 100

96 65.1 49.8 74.6





98 41.4 30.0 45.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#32

1,1,1-Trichloroethane

Concen: 9.748 ppbv

RT: 6.012 min Scan# 1060

Delta R.T. -0.006 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11




Tgt Ion: 97 Resp: 1443296


Ion Ratio Lower Upper

97 100

99 64.9 51.7 77.5

61 53.6 41.1 61.7





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 6.658 min Scan# 1261

Delta R.T. -0.010 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion:114 Resp: 266964

Ion Ratio Lower Upper

114 100

63 25.7 21.8 32.6





88 18.1 15.2 22.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#34

2-Butanone

Concen: 9.182 ppbv

RT: 4.796 min Scan# 682

Delta R.T. -0.003 min

Lab File: VL041298.D




Acq: 4 Jun 2024 9:11


Tgt Ion: 43 Resp: 1247580

Ion Ratio Lower Upper

43 100


72 21.7 17.3 25.9





#35
Carbon Tetrachloride
Concen: 9.986 ppbv
RT: 6.500 min Scan# 1174
Delta R.T. -0.010 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument : MSVOA_L
ClientSampleId : VSTDICCC010



Tgt Ion:117 Resp: 1496089
Ion Ratio Lower Upper
117 100
119 92.5 74.3 111.5
121 31.0 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#36
Benzene
Concen: 8.841 ppbv
RT: 6.378 min Scan# 1174
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 78 Resp: 1980600
Ion Ratio Lower Upper
78 100
77 23.3 18.3 27.5
50 20.0 15.9 23.9





#37
1,2-Dichloroethane
Concen: 9.922 ppbv
RT: 5.816 min Scan# 999
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 62 Resp: 115981
Ion Ratio Lower Upper
62 100
98 6.0 5.0 7.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#38
Trichloroethene
Concen: 9.018 ppbv
RT: 7.295 min Scan# 1459
Delta R.T. -0.013 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11



Tgt Ion:130 Resp: 752906
Ion Ratio Lower Upper
130 100
95 117.5 95.0 142.6
132 95.9 77.0 115.4
97 74.8 59.1 88.7





#39

1,2-Dichloropropane

Concen: 8.877 ppbv

RT: 7.079 min Scan# 1392

Delta R.T. -0.010 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010





Tgt Ion: 63 Resp: 766472

Ion	Ratio	Lower	Upper
63	100		
41	74.9	59.4	89.0
62	67.2	56.8	85.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#40

1,4-Dioxane

Concen: 9.262 ppbv

RT: 7.295 min Scan# 1459




Delta R.T. -0.003 min


Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Tgt Ion: 88 Resp: 285688


Ion	Ratio	Lower	Upper
88	100		
58	134.4	106.1	159.1
43	277.2	219.8	329.6
57	1170.4	936.0	1404.0





#41
Tetrahydrofuran
Concen: 8.279 ppbv
RT: 5.565 min Scan# 91
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion: 42 Resp: 81440
Ion Ratio Lower Upper
42 100
72 37.8 29.4 44.0
71 37.9 29.1 43.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#42
Bromodichloromethane
Concen: 10.055 ppbv
RT: 7.256 min Scan# 1447
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 83 Resp: 1685711
Ion Ratio Lower Upper
83 100
85 63.2 50.6 75.8
127 7.0 6.2 9.4





#43
Methyl Methacrylate
Concen: 8.947 ppbv
RT: 7.478 min Scan# 1516
Delta R.T. -0.010 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion: 69 Resp: 810680
Ion Ratio Lower Upper
69 100
41 157.2 125.2 187.8
100 33.8 26.7 40.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#44
2,2,4-Trimethylpentane
Concen: 8.644 ppbv
RT: 7.333 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Tgt Ion: 57 Resp: 3339096
Ion Ratio Lower Upper
57 100
41 36.8 29.5 44.3
56 31.2 25.3 37.9





#45

t-1,3-Dichloropropene

Concen: 9.110 ppbv

RT: 8.713 min Scan# 1900

Delta R.T. -0.010 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion: 75 Resp: 598199

Ion Ratio Lower Upper

75 100





77 30.9 25.6 38.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#46

cis-1,3-Dichloropropene

Concen: 8.832 ppbv

RT: 8.143 min Scan# 1723

Delta R.T. -0.013 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11




Tgt Ion: 75 Resp: 709965


Ion Ratio Lower Upper

75 100

39 74.0 59.4 89.2

77 31.3 25.9 38.9





#47

1,1,2-Trichloroethane

Concen: 9.548 ppbv

RT: 8.899 min Scan# 1958

Delta R.T. -0.016 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion: 97 Resp: 82514

Ion Ratio Lower Upper

97 100

83 87.1 71.2 106.8

85 53.9 46.0 69.0





99 58.8 49.7 74.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#48

Dibromochloromethane

Concen: 10.198 ppbv

RT: 9.709 min Scan# 2210

Delta R.T. -0.016 min

Lab File: VL041298.D




Acq: 4 Jun 2024 9:11


Tgt Ion:129 Resp: 1404173

Ion Ratio Lower Upper

129 100

127 78.3 39.1 117.5





#49

Bromoform

Concen: 9.891 ppbv

RT: 12.404 min Scan# 3048

Delta R.T. -0.016 min

Lab File: VL041298.D


Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010



Tgt Ion:173 Resp: 115855

Ion Ratio Lower Upper

173 100

175 49.9 39.8 59.8




171 53.2 41.8 62.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#50

4-Methyl-2-Pentanone

Concen: 8.622 ppbv

RT: 8.179 min Scan# 1734

Delta R.T. -0.016 min

Lab File: VL041298.D




Acq: 4 Jun 2024 9:11


Tgt Ion: 43 Resp: 2102376

Ion Ratio Lower Upper

43 100


58 34.6 27.7 41.5





#51
2-Hexanone
Concen: 8.576 ppbv
RT: 9.549 min Scan# 2160
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion: 43 Resp: 1551244
Ion Ratio Lower Upper
43 100
58 46.7 38.4 57.6
98 0.3 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#52
Tetrachloroethene
Concen: 8.988 ppbv
RT: 10.613 min Scan# 2491
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion:164 Resp: 707765
Ion Ratio Lower Upper
164 100
166 124.2 101.1 151.7
129 106.8 81.7 122.5
131 102.6 80.4 120.6





#53
Toluene
Concen: 8.940 ppbv
RT: 9.221 min Scan# 2058
Delta R.T. -0.013 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument : MSVOA_L
ClientSampleId : VSTDICCC010




Tgt Ion: 91 Resp: 2322300
Ion Ratio Lower Upper
91 100
92 61.5 48.2 72.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#54
1,2-Dibromoethane
Concen: 9.626 ppbv
RT: 10.008 min Scan# 2303
Delta R.T. -0.013 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion:107 Resp: 1126770
Ion Ratio Lower Upper
107 100
109 93.8 75.2 112.8





#55
Chlorobenzene-d5
Concen: 10.000 ppbv m
RT: 11.462 min Scan# 2755
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion:117 Resp: 256840
Ion Ratio Lower Upper
117 100
82 65.4 53.6 80.4
119 32.7 27.2 40.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#56
1,1,1,2-Tetrachloroethane
Concen: 9.857 ppbv
RT: 11.507 min Scan# 2769
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion:131 Resp: 1026568
Ion Ratio Lower Upper
131 100
133 94.7 76.3 114.5
119 60.4 49.4 74.2





#57
Chlorobenzene
Concen: 9.286 ppbv
RT: 11.526 min Scan# 2775
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion:112 Resp: 174661
Ion Ratio Lower Upper
112 100
77 65.0 54.9 82.3
114 33.1 28.7 35.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#58
Ethyl Benzene
Concen: 8.878 ppbv
RT: 12.082 min Scan# 2948
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Tgt Ion: 91 Resp: 3126554
Ion Ratio Lower Upper
91 100
106 31.9 24.3 36.5









#59
m/p-Xylene
Concen: 17.931 ppbv m
RT: 12.365 min Scan# 3036
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 91 Resp: 5307470
Ion Ratio Lower Upper
91 100
106 44.6 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#60
o-Xylene
Concen: 8.988 ppbv
RT: 13.053 min Scan# 3250
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 91 Resp: 2493688
Ion Ratio Lower Upper
91 100
106 45.6 22.4 67.3





#61
Styrene
Concen: 9.202 ppbv
RT: 12.892 min Scan# 31
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion:104 Resp: 1348748
Ion Ratio Lower Upper
104 100
78 52.8 42.6 63.8
103 47.4 38.4 57.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#62
Isopropylbenzene
Concen: 9.008 ppbv
RT: 14.021 min Scan# 3551
Delta R.T. -0.022 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion:105 Resp: 3738569
Ion Ratio Lower Upper
105 100
120 25.5 20.2 30.2





#63
1,1,2,2-Tetrachloroethane
Concen: 9.448 ppbv
RT: 13.044 min Scan# 31
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion: 83 Resp: 178675
Ion Ratio Lower Upper
83 100
131 9.4 4.5 13.5
85 65.1 32.7 98.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#64
n-propylbenzene
Concen: 9.610 ppbv
RT: 14.908 min Scan# 3827
Delta R.T. -0.022 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11



Tgt Ion:120 Resp: 995001
Ion Ratio Lower Upper
120 100
91 467.7 384.8 577.2









#65
tert-Butylbenzene
Concen: 8.907 ppbv
RT: 16.082 min Scan# 4192
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010

Tgt Ion: 119 Resp: 331080
Ion Ratio Lower Upper
119 100
91 83.1 67.0 100.4
134 19.6 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#66
Benzyl Chloride
Concen: 6.437 ppbv m
RT: 16.323 min Scan# 4267
Delta R.T. -0.022 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11


Tgt Ion: 91 Resp: 270848
Ion Ratio Lower Upper
91 100
126 16.8 14.6 21.8
128 5.7 4.7 7.1





#67
 sec-Butylbenzene
 Concen: 8.976 ppbv
 RT: 16.629 min Scan# 41
 Delta R.T. -0.019 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010



Tgt Ion:105 Resp: 4696700
 Ion Ratio Lower Upper
 105 100
 134 17.9 13.8 20.8




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
 Supervised By :Mahesh Dadoda 06/05/2024



#68
 1-Bromo-4-Fluorobenzene
 Concen: 10.204 ppbv
 RT: 13.770 min Scan# 3473
 Delta R.T. -0.016 min
 Lab File: VL041298.D
 Acq: 4 Jun 2024 9:11


Tgt Ion: 95 Resp: 1978642
 Ion Ratio Lower Upper
 95 100
 174 65.2 53.6 80.4
 175 5.1 4.2 6.4





#69
p-Isopropyltoluene
Concen: 9.048 ppbv
RT: 16.989 min Scan# 4474
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion:119 Resp: 3901704
Ion Ratio Lower Upper
119 100
134 25.1 20.1 30.1
91 30.1 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#70
n-Butylbenzene
Concen: 9.076 ppbv
RT: 17.876 min Scan# 4750
Delta R.T. -0.022 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Tgt Ion: 91 Resp: 4087151
Ion Ratio Lower Upper
91 100
92 53.6 43.3 64.9
134 22.4 17.8 26.8









#71
2-Chlorotoluene
Concen: 9.195 ppbv
RT: 14.796 min Scan# 3792
Delta R.T. -0.016 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Instrument : MSVOA_L
ClientSampleId : VSTDICCC010

Tgt Ion: 91 Resp: 290554
Ion Ratio Lower Upper
91 100
126 31.1 12.2 48.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024



#72
4-Ethyltoluene
Concen: 9.609 ppbv
RT: 15.188 min Scan# 3914
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Tgt Ion:105 Resp: 3019826
Ion Ratio Lower Upper
105 100
120 29.6 24.7 37.1
91 13.3 10.3 15.5
77 14.5 11.4 17.0





#73

1,3,5-Trimethylbenzene

Concen: 9.187 ppbv

RT: 15.342 min Scan# 3962

Delta R.T. -0.022 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion:105 Resp: 261230

Ion Ratio Lower Upper

105 100





120 47.3 37.0 55.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#74

1,2,4-Trimethylbenzene

Concen: 9.207 ppbv

RT: 16.105 min Scan# 4199

Delta R.T. -0.016 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Tgt Ion:105 Resp: 2859255




Ion Ratio Lower Upper


105 100

120 46.3 37.8 56.6

91 45.3 44.3 66.5


77 20.7 19.0 28.4





#75
1,3-Dichlorobenzene
Concen: 9.672 ppbv
RT: 16.323 min Scan# 4267
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion:146 Resp: 172226
Ion Ratio Lower Upper
146 100
111 46.5 23.7 71.1
148 64.4 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#76
1,4-Dichlorobenzene
Concen: 9.760 ppbv
RT: 16.465 min Scan# 4311
Delta R.T. -0.019 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11



Tgt Ion:146 Resp: 1671398
Ion Ratio Lower Upper
146 100
111 46.6 23.3 69.8
148 65.7 32.6 97.7





#77

1,2-Dichlorobenzene

Concen: 9.365 ppbv

RT: 17.137 min Scan# 4520

Delta R.T. -0.022 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11

Instrument :

MSVOA_L

ClientSampleId :

VSTDICCC010

Tgt Ion:146 Resp: 166691

Ion Ratio Lower Upper

146 100

111 48.6 24.5 73.5





148 64.0 32.6 98.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024

Supervised By :Mahesh Dadoda 06/05/2024



#78

Hexachloro-1,3-Butadiene

Concen: 7.634 ppbv

RT: 22.625 min Scan# 6227

Delta R.T. -0.022 min

Lab File: VL041298.D

Acq: 4 Jun 2024 9:11




Tgt Ion:225 Resp: 1207700


Ion Ratio Lower Upper

225 100

223 63.8 50.9 76.3


227 63.9 52.0 78.0





#79
Naphthalene
Concen: 10.395 ppbv
RT: 21.371 min Scan# 5837
Delta R.T. -0.029 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010




Tgt Ion:128 Resp: 307400
Ion Ratio Lower Upper
128 100
127 13.1 10.8 16.2
129 11.4 8.9 13.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024




#80
Naphthalene,2-methyl-
Concen: 8.865 ppbv
RT: 24.365 min Scan# 6768
Delta R.T. -0.022 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11




Tgt Ion:142 Resp: 1054655
Ion Ratio Lower Upper
142 100
141 85.6 67.9 101.9
115 40.8 31.4 47.0





#81
1,2,4-Trichlorobenzene
Concen: 9.292 ppbv m
RT: 21.130 min Scan# 51
Delta R.T. -0.022 min
Lab File: VL041298.D
Acq: 4 Jun 2024 9:11

Instrument :
MSVOA_L
ClientSampleId :
VSTDICCC010



Tgt Ion:180 Resp: 1450918
Ion Ratio Lower Upper
180 100
182 96.3 77.2 115.8
184 0.0 24.6 36.8

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/05/2024
Supervised By :Mahesh Dadoda 06/05/2024

