

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060524\
 Data File : VL041334.D
 Acq On : 5 Jun 2024 19:33
 Operator : SY/MD
 Sample : P2716-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/06/2024
 Supervised By :Mahesh Dadoda 06/06/2024

Quant Time: Jun 06 04:42:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL060424AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Jun 05 04:42:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.214	49	935398	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.674	114	2462340	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.481	117	2208706	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.786	95	1723589	10.107	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
9) Propene	2.735	41	11998m	0.228	ppbv	
13) Ethanol	3.288	45	3444m	0.251	ppbv	
15) Acetone	3.517	43	43845	0.411	ppbv #	90
20) Methylene Chloride	3.963	84	33036	0.851	ppbv	91
26) Hexane	5.234	57	44192	0.465	ppbv #	42
34) 2-Butanone	4.835	43	24847m	0.211	ppbv	
52) Tetrachloroethene	10.629	164	3876m	0.056	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060524\
Data File : VL041334.D
Acq On : 5 Jun 2024 19:33
Operator : SY/MD
Sample : P2716-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Quant Time: Jun 06 04:42:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL060424AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Jun 05 04:42:10 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 06/06/2024
Supervised By : Mahesh Dadoda 06/06/2024

