

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060719\
Data File : VL033884.D
Acq On : 7 Jun 2019 11:29
Operator : SY/AP
Sample : QC052319 CAN 10262
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC052319 CAN 10262

Quant Time: Jun 09 21:48:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.74	49	857540	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.27	114	1989558	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.22	117	2028688	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.56	95	1663965	9.99	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060719\
Data File : VL033884.D
Acq On : 7 Jun 2019 11:29
Operator : SY/AP
Sample : QC052319 CAN 10262
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC052319 CAN 10262

Quant Time: Jun 09 21:48:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.74 min Scan# 873

Delta R.T. -0.01 min

Lab File: VL033884.D

Acq: 7 Jun 2019 11:29

Instrument :

MSVOA_L

ClientSampleId :

QC052319 CAN 10262


Tot Ion: 49 Resp: 857540

Ion Ratio Lower Upper

49 100

128 50.7 25.5 76.5




130 65.7 32.5 97.4



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.27 min Scan# 1348

Delta R.T. -0.01 min

Lab File: VL033884.D

Acq: 7 Jun 2019 11:29


Tgt Ion: 114 Resp: 1989558

Ion Ratio Lower Upper

114 100

63 27.2 21.1 31.7



88 18.2 14.2 21.4




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.22 min Scan# 2886
Delta R.T. -0.01 min
Lab File: VL033884.D
Acq: 7 Jun 2019 11:29

Instrument :
MSVOA_L
ClientSampleId :
QC052319 CAN 10262

Tot Ion:117 Resp: 2028688
Ion Ratio Lower Upper
117 100
82 68.2 50.2 75.2
119 32.5 24.8 37.2



#68
1-Bromo-4-Fluorobenzene
Concen: 9.986 ppbv
RT: 14.56 min Scan# 3614
Delta R.T. -0.01 min
Lab File: VL033884.D
Acq: 7 Jun 2019 11:29

Tgt Ion: 95 Resp: 1663965
Ion Ratio Lower Upper
95 100
174 66.2 54.0 81.0
175 4.8 3.8 5.8

