

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061322\
 Data File : VL039266.D
 Acq On : 14 Jun 2022 00:40
 Operator : SY/AP
 Sample : N3306-03DL 5X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/14/2022
 Supervised By :Mahesh Dadoda 06/16/2022

Quant Time: Jun 14 01:35:57 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 01 01:22:18 2022

QLast Update : Wed Jun 01 01:22:18 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	1258014	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.176	114	2871692	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.085	117	2407087	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.417	95	2012671	10.114	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
3) Chlorodifluoromethane	2.989	51	22417	0.140	ppbv #	80
9) Propene	3.031	41	65343m	0.936	ppbv	
13) Ethanol	3.607	45	114531m	30.090	ppbv	
15) Acetone	3.858	43	452578	3.454	ppbv	91
19) Isopropyl Alcohol	3.973	45	334256	5.438	ppbv #	89
20) Methylene Chloride	4.343	84	62878	1.139	ppbv	85
26) Hexane	5.693	57	48891	0.377	ppbv #	63
52) Tetrachloroethene	11.217	164	16399m	0.180	ppbv	
53) Toluene	9.812	91	34767m	0.130	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061322\
Data File : VL039266.D
Acq On : 14 Jun 2022 00:40
Operator : SY/AP
Sample : N3306-03DL 5X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/14/2022
Supervised By :Mahesh Dadoda 06/16/2022

Quant Time: Jun 14 01:35:57 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

