

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061322\
 Data File : VL039268.D
 Acq On : 14 Jun 2022 1:59
 Operator : SY/AP
 Sample : N3306-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/14/2022
 Supervised By :Mahesh Dadoda 06/16/2022

Quant Time: Jun 14 02:55:32 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 0

QLast Update : Wed Jun 01 01:22:18 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	1154463	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.176	114	2613225	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.089	117	2171479	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.416	95	1876204	10.451	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.500%
Target Compounds						
					Qvalue	
9) Propene	3.028	41	27422	0.428	ppbv	84
13) Ethanol	3.616	45	45306m	12.970	ppbv	
15) Acetone	3.861	43	192030	1.597	ppbv	100
19) Isopropyl Alcohol	3.980	45	167640m	2.972	ppbv	
20) Methylene Chloride	4.346	84	34256	0.676	ppbv #	93
52) Tetrachloroethene	11.227	164	7206m	0.087	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061322\
Data File : VL039268.D
Acq On : 14 Jun 2022 1:59
Operator : SY/AP
Sample : N3306-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-1DL

Quant Time: Jun 14 02:55:32 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/14/2022
Supervised By :Mahesh Dadoda 06/16/2022

