

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061422\
 Data File : VL039287.D
 Acq On : 14 Jun 2022 19:25
 Operator : SY/AP
 Sample : N3306-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/16/2022
 Supervised By :Mahesh Dadoda 06/16/2022

Quant Time: Jun 15 03:20:27 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 01 01:22:18 2022

QLast Update : Wed Jun 01 01:22:18 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.674	49	1329271	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.179	114	3721607	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.092	117	3260601	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.416	95	2501326	9.279	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.800%
Target Compounds						
						Qvalue
9) Propene	3.025	41	23085m	0.313	ppbv	
13) Ethanol	3.607	45	31296m	7.781	ppbv	
15) Acetone	3.854	43	286948	2.072	ppbv	96
19) Isopropyl Alcohol	3.976	45	34809m	0.536	ppbv	
20) Methylene Chloride	4.353	84	77604	1.330	ppbv	98
26) Hexane	5.700	57	61819	0.451	ppbv #	39
34) 2-Butanone	5.256	43	35019m	0.173	ppbv	
52) Tetrachloroethene	11.220	164	14785m	0.125	ppbv	
53) Toluene	9.809	91	34203m	0.098	ppbv	
59) m/p-Xylene	12.970	91	71478m	0.181	ppbv	
60) o-Xylene	13.696	91	38792m	0.104	ppbv	
74) 1,2,4-Trimethylbenzene	16.767	105	90718m	0.207	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061422\
Data File : VL039287.D
Acq On : 14 Jun 2022 19:25
Operator : SY/AP
Sample : N3306-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-1DL

Quant Time: Jun 15 03:20:27 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/16/2022
Supervised By :Mahesh Dadoda 06/16/2022

