

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL061920\
 Data File : VL035100.D
 Acq On : 19 Jun 2020 18:27
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 6/22/2020 2:09:04 PM

Quant Time: Jun 20 02:02:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun 20 02:02:44 2020
 QLast Update : Sat Jun 20 01:15:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1320073	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3041578	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	2869149	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2473576	10.69	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.21	62	7971m	0.095	ppbv	
32) 1,1,1-Trichloroethane	6.50	97	18633m	0.099	ppbv	
35) Carbon Tetrachloride	7.01	117	18141m	0.095	ppbv	
38) Trichloroethene	7.83	130	9093m	0.093	ppbv	
52) Tetrachloroethene	11.24	164	8888m	0.100	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.71	83	23294m	0.097	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL061920\
Data File : VL035100.D
Acq On : 19 Jun 2020 18:27
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
6/22/2020 2:09:04 PM

Quant Time: Jun 20 02:02:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun
QLast Update : Sat Jun 20 01:15:38 2020
Response via : Initial Calibration

