

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062221\
 Data File : VL037238.D
 Acq On : 23 Jun 2021 9:07
 Operator : SY/AP
 Sample : M2797-01DL 200X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

MMDadoda
 6/25/2021 2:21:22 PM

Quant Time: Jun 24 09:53:03 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 1
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1581109	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	4562603	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.07	117	3987244	10.00	ppbv	-0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2851595	9.44	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	2.98	51	152116	0.495	ppbv	100
9) Propene	3.00	41	148454	1.510	ppbv	99
15) Acetone	3.84	43	779888m	4.374	ppbv	
34) 2-Butanone	5.24	43	1023968	5.771	ppbv	99
36) Benzene	6.87	78	593606	1.609	ppbv	99
52) Tetrachloroethene	11.21	164	19955m	0.116	ppbv	
53) Toluene	9.79	91	188433m	0.428	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL062221\
Data File : VL037238.D
Acq On : 23 Jun 2021 9:07
Operator : SY/AP
Sample : M2797-01DL 200X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

MMDadoda
6/25/2021 2:21:22 PM

Quant Time: Jun 24 09:53:03 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

