

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062221\
 Data File : VL037240.D
 Acq On : 23 Jun 2021 10:42
 Operator : SY/AP
 Sample : M2797-03DL 200X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV3DL

Quant Time: Jun 24 09:54:48 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Jun 16 14:06:32 2021
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1537259	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	4439757	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.07	117	3899365	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2876867	9.74	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	2.98	51	54657	0.183	ppbv	98
9) Propene	3.00	41	93724	0.981	ppbv	97
15) Acetone	3.85	43	509288	2.938	ppbv #	85
34) 2-Butanone	5.25	43	645296	3.737	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062221\
Data File : VL037240.D
Acq On : 23 Jun 2021 10:42
Operator : SY/AP
Sample : M2797-03DL 200X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV3DL

Quant Time: Jun 24 09:54:48 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 16 14:06:32 2021
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

