

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062221\
 Data File : VL037221.D
 Acq On : 22 Jun 2021 17:48
 Operator : SY/AP
 Sample : QC062121 CAN 10792
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062121 CAN 10792

Quant Time: Jun 23 12:21:31 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 16 14:06:32 2021
 QLast Update : Wed Jun 16 14:06:32 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1548026	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	4511349	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3859091	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2934752	10.04	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%


Target Compounds	Ovalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL062221\
Data File : VL037221.D
Acq On : 22 Jun 2021 17:48
Operator : SY/AP
Sample : OC062121 CAN 10792
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062121 CAN 10792

Quant Time: Jun 23 12:21:31 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 16 14:06:32 2021
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.68 min Scan# 901

Delta R.T. 0.01 min

Lab File: VL037221.D

Acq: 22 Jun 2021 17:48

Instrument :

MSVOA_L

ClientSampleId :

QC062121 CAN 10792


Tot Ion: 49 Resp: 1548026

Ion Ratio Lower Upper

49 100

128 56.1 29.3 88.0




130 72.7 35.9 107.8



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.19 min Scan# 1369

Delta R.T. 0.00 min

Lab File: VL037221.D

Acq: 22 Jun 2021 17:48


Tgt Ion: 114 Resp: 4511349

Ion Ratio Lower Upper

114 100

63 24.7 19.6 29.4



88 18.1 14.2 21.2




Abundance Ion 114.10 (113.80 to 114.80): VL0

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.10 min Scan# 2897
Delta R.T. 0.01 min
Lab File: VL037221.D
Acq: 22 Jun 2021 17:48

Instrument :
MSVOA_L
ClientSampleId :
QC062121 CAN 10792

Tot Ion:117 Resp: 3859091
Ion Ratio Lower Upper
117 100
82 66.1 51.7 77.5
119 34.8 47.6 71.4#



#68
1-Bromo-4-Fluorobenzene
Concen: 10.036 ppbv
RT: 14.43 min Scan# 3622
Delta R.T. 0.01 min
Lab File: VL037221.D
Acq: 22 Jun 2021 17:48

Tgt Ion: 95 Resp: 2934752
Ion Ratio Lower Upper
95 100
174 73.1 57.0 85.4
175 5.2 4.4 6.6

