

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062222\
 Data File : VL039331.D
 Acq On : 22 Jun 2022 22:10
 Operator : SY/AP
 Sample : QC062122 CAN 10158
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062122 CAN 10158

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/24/2022
 Supervised By : Mahesh Dadoda 06/28/2022

Quant Time: Jun 23 07:30:50 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 23 07:30:50 2022
 QLast Update : Wed Jun 01 01:22:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1089816	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	7.169	114	3093693	10.000	ppbv	-0.01
55) Chlorobenzene-d5	12.079	117	2673123	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.407	95	2070492	9.369	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.700%
Target Compounds						
20) Methylene Chloride	4.343	84	39054	0.816	ppbv	90
26) Hexane	5.684	57	25884m	0.230	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062222\
Data File : VL039331.D
Acq On : 22 Jun 2022 22:10
Operator : SY/AP
Sample : QC062122 CAN 10158
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062122 CAN 10158

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 06/24/2022
Supervised By : Mahesh Dadoda 06/28/2022

Quant Time: Jun 23 07:30:50 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

