

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062419\
Data File : VL033934.D
Acq On : 24 Jun 2019 13:17
Operator : SY/AP
Sample : QC060619 CAN 10729
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC060619 CAN 10729

Quant Time: Jun 25 03:53:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.74	49	1016528	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.26	114	2245419	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.21	117	2239764	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.54	95	1826399	9.93	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL062419\
Data File : VL033934.D
Acq On : 24 Jun 2019 13:17
Operator : SY/AP
Sample : OC060619 CAN 10729
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC060619 CAN 10729

Quant Time: Jun 25 03:53:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration

